

COMPREHENSIVE SERVICES

We offer competitive repair and calibration services, as well as easily accessible documentation and free downloadable resources.

SELL YOUR SURPLUS

We buy new, used, decommissioned, and surplus parts from every NI series. We work out the best solution to suit your individual needs.

 Sell For Cash  Get Credit  Receive a Trade-In Deal

OBSOLETE NI HARDWARE IN STOCK & READY TO SHIP

We stock **New**, **New Surplus**, **Refurbished**, and **Reconditioned** NI Hardware.



Bridging the gap between the manufacturer and your legacy test system.

 1-800-915-6216

 www.apexwaves.com

 sales@apexwaves.com

All trademarks, brands, and brand names are the property of their respective owners.

Request a Quote

 **CLICK HERE**

NI-6585B

GETTING STARTED GUIDE

NI 6585/6585B

Low-Voltage Differential Adapter Module



Note Before you begin, complete the software and hardware installation instructions in your FlexRIO FPGA getting started guide or controller for FlexRIO getting started guide.



Caution Using the NI 6585/6585B in a manner not described in this document may impair the protection the NI 6585/6585B provides.

The NI 6585/6585B is a low-voltage differential (LVDS) adapter module designed to work in conjunction with FlexRIO FPGA modules and controllers for FlexRIO.

This document explains how to install and configure the NI 6585/6585B.



Note *NI 6585R* refers to the combination of your NI 6585/6585B adapter module and either a FlexRIO FPGA module or a controller for FlexRIO. *NI 6585/6585B* refers to your adapter module only.

Contents

Electromagnetic Compatibility Guidelines.....	1
FlexRIO Documentation.....	2
Verifying the System Requirements.....	4
Compatibility with FPGA Modules and Controllers for FlexRIO.....	4
Unpacking.....	4
Preparing the Environment.....	5
Installing the NI 6585/6585B.....	5
Confirming that Measurement & Automation Explorer (MAX) Recognizes the Device.....	5
Front Panel and Connector Pinouts.....	6
Pinout and Signal Information.....	7
Component-Level Intellectual Property (CLIP).....	11
NI 6585/6585B CLIP.....	11
Worldwide Support and Services.....	11
Where to Go Next.....	12

Electromagnetic Compatibility Guidelines

This product was tested and complies with the regulatory requirements and limits for electromagnetic compatibility (EMC) stated in the product specifications. These requirements and limits are designed to provide reasonable protection against harmful interference when the product is operated in the intended operational electromagnetic environment.

This product is intended for use in industrial locations. However, harmful interference may occur in some installations, when the product is connected to a peripheral device or test object, or if the product is used in residential or commercial areas. To minimize interference with radio and television reception and prevent unacceptable performance degradation, install and use this product in strict accordance with the instructions in the product documentation.

Furthermore, any modifications to the product not expressly approved by National Instruments could void your authority to operate it under your local regulatory rules.



Caution To ensure the specified EMC performance, operate this product only with shielded cables and accessories.



Caution To ensure the specified EMC performance, you must install PXI EMC Filler Panels, National Instruments part number 778700-01, in all open chassis slots. You can order a kit of six single slot EMC Filler Panels directly from NI by visiting ni.com.



Caution To ensure the specified EMC performance, the length of all I/O cables must be no longer than 30 m (100 ft).



Caution Using the NI 6585/6585B in a manner not described in this document may impair the protection the NI 6585/6585B provides.

FlexRIO Documentation

Table 1. FlexRIO Documentation Locations and Descriptions

Document	Location	Description
Getting started guide for your FlexRIO FPGA module or controller for FlexRIO	Available from the Start menu and at ni.com/manuals .	Contains installation instructions for your FlexRIO system.
Specifications document for your FlexRIO FPGA module or controller for FlexRIO	Available from the Start menu and at ni.com/manuals .	Contains specifications for your FlexRIO FPGA module or controller for FlexRIO.
Getting started guide for your adapter module	Available from the Start menu and at ni.com/manuals .	Contains signal information, examples, and CLIP details for your adapter module.
Specifications document for your adapter module	Available from the Start menu and at ni.com/manuals .	Contains specifications for your adapter module.

Table 1. FlexRIO Documentation Locations and Descriptions (Continued)

Document	Location	Description
<i>LabVIEW FPGA Module Help</i>	Embedded in <i>LabVIEW Help</i> and at ni.com/manuals .	Contains information about the basic functionality of the LabVIEW FPGA Module.
<i>Real-Time Module Help</i>	Embedded in <i>LabVIEW Help</i> and at ni.com/manuals .	Contains information about real-time programming concepts, step-by-step instructions for using LabVIEW with the Real-Time Module, reference information about Real-Time Module VIs and functions, and information about LabVIEW features on real-time operating systems.
<i>FlexRIO Help</i>	Available from the Start menu and at ni.com/manuals .	Contains information about the FPGA module front panel connectors and I/O, controller for FlexRIO front panel connectors and I/O, programming instructions, and adapter module component-level IP (CLIP).
<i>FlexRIO Adapter Module Development Kit User Manual</i>	Available from the Start menu at Start»All Programs»National Instruments»NI FlexRIO»NI FlexRIO Adapter Module Development Kit»Documentation .	Contains information about how to create custom adapter modules for use with FlexRIO FPGA modules.
LabVIEW Examples	Available in NI Example Finder. In LabVIEW, click Help»Find Examples»Hardware Input and Output»FlexRIO .	Contains examples of how to run FPGA VIs and Host VIs on your device.
IPNet	Located at ni.com/ipnet .	Contains LabVIEW FPGA functions and intellectual property to share.
FlexRIO product page	Located at ni.com/flexrio .	Contains product information and data sheets for FlexRIO devices.

Verifying the System Requirements

To use the NI 6585/6585B, your system must meet certain requirements. For more information about minimum system requirements, recommended system, and supported application development environments (ADEs), refer to the readme, which is available on the software media or online at ni.com/updates.

Compatibility with FPGA Modules and Controllers for FlexRIO

The NI 6585B variant is compatible with all FlexRIO FPGA modules and controllers for FlexRIO; however, the NI 6585 variant is only compatible with NI PXI-795xR and NI PXIe-796xR modules. The following table lists each adapter module's compatibility with FlexRIO FPGA modules and controllers for FlexRIO.

Table 2. Adapter Module Compatibility

	NI PXI-795xR	NI PXIe-796xR	NI PXIe-797xR	NI-793xR
NI 6585	Yes	Yes	No	No
NI 6585B	Yes	Yes	Yes	Yes

Unpacking



Caution To prevent ESD from damaging the devices, ground yourself using a grounding strap or by holding a grounded object, such as your computer chassis.

1. Touch the antistatic package to a metal part of the computer chassis.
2. Remove each module from the package and inspect it for loose components or any other sign of damage.



Caution Never touch the exposed pins of connectors.



Note Do not install a device if it appears damaged in any way.

3. Unpack any other items and documentation from the kit.

Store the devices in the antistatic package when they are not in use.

Preparing the Environment

Ensure that the environment you are using the NI 6585/6585B in meets the following specifications.

Operating temperature (IEC 60068-2-1, IEC 60068-2-2)	0 °C to 55 °C
Operating humidity (IEC 60068-2-56)	10% to 90% RH, noncondensing
Pollution Degree	2
Maximum altitude	2,000 m at 25 °C ambient temperature

Indoor use only.



Note Refer to the *NI 6585/6585B Specifications* at ni.com/manuals for complete specifications.



Caution Clean the hardware with a soft, nonmetallic brush. Make sure that the hardware is completely dry and free from contaminants before returning it to service.

Installing the NI 6585/6585B

Refer to the getting started guide for your FlexRIO FPGA module or Controller for FlexRIO for instructions about how to install your FlexRIO system, including the NI 6585/6585B.

Confirming that Measurement & Automation Explorer (MAX) Recognizes the Device

Use Measurement & Automation Explorer (MAX) to configure your NI hardware. MAX informs other programs about which devices reside in the system and how they are configured. MAX is automatically installed with FlexRIO Support.

1. Launch MAX by navigating to **Start»All Programs»National Instruments»NI MAX** or by clicking the NI MAX desktop icon.
2. In the Configuration pane, double-click **Devices and Interfaces** to see the list of installed devices. Installed devices appear under the name of their associated chassis.
3. (PXI and PXI Express devices only) Expand your **Chassis** tree item. MAX lists all devices installed in the chassis. Your default device names may vary.




Note If you do not see your hardware listed, press <F5> to refresh the list of installed devices. If the device is still not listed, power off the system, ensure the device is correctly installed, and restart.

4. (Controllers for FlexRIO only) Your device appears under the **Remote Devices** section.

Front Panel and Connector Pinouts

The following figure shows the front panel connectors for the NI 6585/6585B.

Figure 1. NI 6585/6585B Connector Pin Assignments



Caution To avoid permanent damage to the NI 6585/6585B, disconnect all signals connected to the NI 6585/6585B before powering down the module, and connect signals only after the adapter module has been powered on by the FlexRIO FPGA module or controller for FlexRIO.



Caution Connections that exceed any of the maximum ratings of any connector on the NI 6585/6585B can damage the device and the chassis. NI is not liable for any damage resulting from such connections.

Pinout and Signal Information

The following table contains pin location and signal information for the NI 6585/6585B.



Note DIO <0..15>+/-, PFI <0..4>+/-, and DDC CLOCK OUT appear on both connectors, DDCA and DDCB.

Table 3. DDC Connector Pins

Signal Name	Pin(s)	Signal Type	Signal Description
GLOBAL CLOCK 0+	26 on DDCA	Clock	Input terminal for the external clock source, which can be used for dynamic acquisition or generation.
GLOBAL CLOCK 0-	27 on DDCA		
GLOBAL CLOCK 1+	26 on DDCB		
GLOBAL CLOCK 1-	27 on DDCB		
DIO <0..15>+	2, 5, 8, 11, 14, 17, 20, 23, 37, 40, 43, 46, 49, 52, 55, 58	Data	Positive terminal of LVDS digital I/O data channels 0 through 15.
DIO <0..15>-	3, 6, 9, 12, 15, 18, 21, 24, 38, 41, 44, 47, 50, 53, 56, 59	Data	Negative terminal of LVDS digital I/O data channels 0 through 15.
DDC CLOCK OUT+/PFI 0-	61	Clock/Data	Positive output terminal for the exported Sample clock/Positive Programmable Function Interface terminal 0
DDC CLOCK OUT+/PFI 0-	62	Clock/Data	Negative output terminal for the exported Sample clock/Negative Programmable Function Interface terminal 0
PFI <1..4>+	29, 64, 32, 67	Data	Positive terminal of the Programmable Function Interface LVDS digital I/O channels 1 through 4.

Table 3. DDC Connector Pins (Continued)

Signal Name	Pin(s)	Signal Type	Signal Description
PFI <1..4>-	30, 65, 33, 68	Data	Negative terminal of the Programmable Function Interface LVDS digital I/O channels 1 through 4.
GND	1, 4, 7, 10, 13, 16, 19, 22, 25, 28, 31, 34, 35, 36, 39, 42, 45, 48, 51, 54, 57, 60, 63, 66	Ground	Ground reference for signals.

The following table lists the NI 6585/6585B connector signals and corresponding FlexRIO FPGA module signals necessary for designing a custom CLIP.

Table 4. NI 6585/6585B Connector Signals and FlexRIO FPGA Module Signals

NI 6585/6585B		FlexRIO FPGA Module		
Connector	Signal Name	GPIO Out	GPIO Out Enable	GPIO In
DDCA	GLOBAL CLOCK 0	—	—	GC1k_SE
	DIO 0+/-	GPIO_30_n	GPIO_31	GPIO_31_n
	DIO 1+/-	GPIO_12_n	GPIO_13	GPIO_13_n
	DIO 2+/-	GPIO_32	GPIO_32_n	GPIO_15_n
	DIO 3+/-	GPIO_14	GPIO_14_n	GPIO_15
	DIO 4+/-	GPIO_27_n	GPIO_28	GPIO_28_n
	DIO 5+/-	GPIO_9_n	GPIO_10	GPIO_10_n
	DIO 6+/-	GPIO_29	GPIO_29_n	GPIO_30
	DIO 7+/-	GPIO_11	GPIO_11_n	GPIO_12
	DIO 8+/-	GPIO_22	GPIO_22_n	GPIO_23
	DIO 9+/-	GPIO_6_n	GPIO_7	GPIO_7_n
	DIO 10+/-	GPIO_26	GPIO_26_n	GPIO_27
	DIO 11+/-	GPIO_8	GPIO_8_n	GPIO_9
	DIO 12+/-	GPIO_19	GPIO_19_n	GPIO_20
	DIO 13+/-	GPIO_3_n	GPIO_4	GPIO_4_n
	DIO 14+/-	GPIO_20_n	GPIO_21	GPIO_21_n
	DIO 15+/-	GPIO_5	GPIO_5_n	GPIO_6
	DDC CLOCK OUT+/PFI 0+ DDC CLOCK OUT-/PFI 0-	GPIO_0	GPIO_0_n	GPIO_23_n
	PFI 1+/-	GPIO_1	GPIO_1_n	GPIO_24
	PFI 2+/-	GPIO_16	GPIO_16_n	GPIO_24_n
	PFI 3+/-	GPIO_17	GPIO_17_n	GPIO_25
	PFI 4+/-	GPIO_18	GPIO_18_n	GPIO_25_n

Table 5. NI 6585/6585B Connector Signals and FlexRIO FPGA Module Signals
(Continued)

NI 6585/6585B		FlexRIO FPGA Module		
Connector	Signal Name	GPIO Out	GPIO Out Enable	GPIO In
DDCB	GLOBAL CLOCK 1	—	—	GClk_LVDS
	DIO 0+/-	GPIO_46	GPIO_46_n	GPIO_47
	DIO 1+/-	GPIO_62_n	GPIO_63	GPIO_63_n
	DIO 2+/-	GPIO_47_n	GPIO_48	GPIO_48_n
	DIO 3+/-	GPIO_64	GPIO_64_n	GPIO_65
	DIO 4+/-	GPIO_43	GPIO_43_n	GPIO_44
	DIO 5+/-	GPIO_59_n	GPIO_60	GPIO_60_n
	DIO 6+/-	GPIO_44_n	GPIO_45	GPIO_45_n
	DIO 7+/-	GPIO_61	GPIO_61_n	GPIO_62
	DIO 8+/-	GPIO_39	GPIO_39_n	GPIO_40
	DIO 9+/-	GPIO_54	GPIO_54_n	GPIO_55
	DIO 10+/-	GPIO_40_n	GPIO_41	GPIO_41_n
	DIO 11+/-	GPIO_55_n	GPIO_56	GPIO_56_n
	DIO 12+/-	GPIO_36	GPIO_36_n	GPIO_37
	DIO 13+/-	GPIO_51	GPIO_51_n	GPIO_52
	DIO 14+/-	GPIO_37_n	GPIO_38	GPIO_38_n
	DIO 15+/-	GPIO_52_n	GPIO_53	GPIO_53_n
	DDC CLOCK OUT+/PFI 0+ DDC CLOCK OUT-/PFI 0-	GPIO_49	GPIO_49_n	GPIO_57
	PFI 1+/-	GPIO_50	GPIO_50_n	GPIO_57_n
	PFI 2+/-	GPIO_33	GPIO_33_n	GPIO_58
	PFI 3+/-	GPIO_34	GPIO_34_n	GPIO_58_n
	PFI 4+/-	GPIO_35	GPIO_35_n	GPIO_59

Component-Level Intellectual Property (CLIP)

The LabVIEW FPGA Module includes component-level intellectual property (CLIP) for HDL IP integration. FlexRIO devices support two types of CLIP: user-defined and socketed.

- *User-defined CLIP* allows you to insert HDL IP into an FPGA target, enabling VHDL code to communicate directly with an FPGA VI.
- *Socketed CLIP* provides the same IP integration of the user-defined CLIP, but it also allows the CLIP to communicate directly with circuitry external to the FPGA. Adapter module socketed CLIP allows your IP to communicate directly with both the FPGA VI and the external adapter module connector interface.

The FlexRIO adapter module ships with socketed CLIP items that add module I/O to the LabVIEW project.

NI 6585/6585B CLIP

The NI 6585/6585B ships with socketed CLIP that adds module I/O to the LabVIEW project.

The NI-developed NI 6585/6585B CLIP are as follows:

- **NI 6585 Basic Channel**—Provides read/write access to all low-voltage differential signal (LVDS) channels using a simple channel-based interface. Each I/O line has a write enable signal. This CLIP provides a clock signal for export on each connector. The clock inputs from the NI 6585 are passed to LabVIEW FPGA for use in the FPGA VI. This CLIP also allows for individual clock output inversion.
- **NI 6585 Basic Connector CLIP**—Provides read/write access to all low-voltage differential signal (LVDS) lines on each connector, where the lines are grouped per connector. The individual data lines for each connector are accessed using a U16 data type in LabVIEW FPGA. Each I/O line has a write enable signal. This CLIP also allows for individual clock output inversion.
- **NI 6585 DDR Connector CLIP**—Provides read/write access to all low-voltage differential signal (LVDS) lines on each connector, where the lines are grouped per connector. The individual datalines for each connector are accessed using a U16 data type in LabVIEW FPGA. Data from each edge of the clock is presented as “rising” and “falling.” Each I/O line has a write enable signal. This CLIP also allows for individual clock output inversion.

Related Information

Refer to the [FlexRIO Help](#) for information about FlexRIO CLIP, configuring the NI 6585/6585B with a socketed CLIP, and a list of available socketed CLIP and provided signals.

Worldwide Support and Services

The NI website is your complete resource for technical support. At ni.com/support, you have access to everything from troubleshooting and application development self-help resources to email and phone assistance from NI Application Engineers.

Visit ni.com/services for NI Factory Installation Services, repairs, extended warranty, and other services.


Visit ni.com/register to register your NI product. Product registration facilitates technical support and ensures that you receive important information updates from NI.

A Declaration of Conformity (DoC) is our claim of compliance with the Council of the European Communities using the manufacturer's declaration of conformity. This system affords the user protection for electromagnetic compatibility (EMC) and product safety. You can obtain the DoC for your product by visiting ni.com/certification. If your product supports calibration, you can obtain the calibration certificate for your product at ni.com/calibration.

NI corporate headquarters is located at 11500 North Mopac Expressway, Austin, Texas, 78759-3504. NI also has offices located around the world. For telephone support in the United States, create your service request at ni.com/support or dial 1 866 ASK MYNI (275 6964). For telephone support outside the United States, visit the *Worldwide Offices* section of ni.com/niglobal to access the branch office websites, which provide up-to-date contact information, support phone numbers, email addresses, and current events.

Where to Go Next

Refer to the following figure for information about other product tasks and associated resources for those tasks.



Refer to the *NI Trademarks and Logo Guidelines* at ni.com/trademarks for information on NI trademarks. Other product and company names mentioned herein are trademarks or trade names of their respective companies. For patents covering NI products/technology, refer to the appropriate location: **Help»Patents** in your software, the `patents.txt` file on your media, or the *National Instruments Patent Notice* at ni.com/patents. You can find information about end-user license agreements (EULAs) and third-party legal notices in the readme file for your NI product. Refer to the *Export Compliance Information* at ni.com/legal/export-compliance for the NI global trade compliance policy and how to obtain relevant HTS codes, ECCNs, and other import/export data. NI MAKES NO EXPRESS OR IMPLIED WARRANTIES AS TO THE ACCURACY OF THE INFORMATION CONTAINED HEREIN AND SHALL NOT BE LIABLE FOR ANY ERRORS. U.S. Government Customers: The data contained in this manual was developed at private expense and is subject to the applicable limited rights and restricted data rights as set forth in FAR 52.227-14, DFAR 252.227-7014, and DFAR 252.227-7015.

© 2015 National Instruments. All rights reserved.