COMPREHENSIVE SERVICES

We offer competitive repair and calibration services, as well as easily accessible documentation and free downloadable resources.

SELL YOUR SURPLUS

We buy new, used, decommissioned, and surplus parts from every NI series. We work out the best solution to suit your individual needs.

Sell For Cash Get Credit Receive a Trade-In Deal

OBSOLETE NI HARDWARE IN STOCK & READY TO SHIP

We stock New, New Surplus, Refurbished, and Reconditioned NI Hardware.



Bridging the gap between the manufacturer and your legacy test system.

0

1-800-915-6216



www.apexwaves.com

sales@apexwaves.com

All trademarks, brands, and brand names are the property of their respective owners.

Request a Quote



PXI-7352



Board Assembly Part Number(s)

Part Number	Description
190975E-02L or later	PCI-7352
190975E-04L or later	PCI-7354
190975E-06L or later	PCI-7356
190975E-08L or later	PCI-7358
190974G-02L or later	PXI-7352
190974G-04L or later	PXI-7354
190974G-06L or later	PXI-7356
190974G-08L or later	PXI-7358

Manufacturer: National Instruments

Volatile Memory

Type	Size	User Accessible/ System Accessible ¹	Battery Backup?	Purpose	Method of Clearing ²
FPGA	1,000,000	Gates No/Yes	No	Encoder, Limits, DACs, ADCs, DSP Interface	Cycle power
FPGA	100,000 Ga	ntes No/Yes	No	Microprocessor to Host Interface, Digital I/O	Cycle Power
CPLD	1,250 Gate	s No/Yes	No	FPGA Configuration	
SRAM	512 KB x2	Yes/Yes	No	Used by the microprocesso Also stores onboard variab which are user accessible	2
SRAM	160 KB	No/Yes	No	Integrated RAM for DSP	Cycle Power
Non-Volatil	e Memory				
Type	Size	User Accessible/ System Accessible	Battery Backup?	Purpose	Method of Clearing
EEPROM	8 KB	No/Yes	No	PCI Configuration	None available to user
Flash	2 MB	Yes/Yes	No	Stores onboard programs, FPGA bitstreams, DSP initialization, microprocessor boot image, buffers, and user	Flash memory pointers can be cleared by using the memory management function or by using MAX.

defaults. The user can modify the onboard programs, buffers,

and user default values

EDITION DATE January 2016 NI PART NUMBER 375919A-01 The flash is not actually

accessible from any API.


cleared, but it is not

¹ Items are designated **No** for the following reason(s):

a) Hardware changes or a unique software tool from National Instruments are required to modify contents of the memory listed.

b) Hardware-modifying software tools are not distributed to customers for any personal access or customization, also known as non-normal use.

The designation None Available to User indicates that the ability to clear this memory is not available to the user under normal operation. The utilities required to clear the memory are not distributed by National Instruments to customers for normal use.





Media Storage

User Accessible/ Battery
Size System Accessible Backup? Purpose Method of Clearing

NONE

Type

Terms and Definitions

User Accessible Allows the user to directly write or modify the contents of the memory during normal instrument operation.

System Accessible Does not allow the user to access or modify the memory during normal instrument operation. However, system accessible memory may be accessed or modified by background processes. This can be something that is not deliberate by the user and can be a background driver implementation, such as storing application information in RAM to increase speed of use.

Cycle Power The process of completely removing power from the device and its components. This process includes a complete shutdown of the PC and/or chassis containing the device; a reboot is not sufficient for the completion of this process.

Volatile Memory Requires power to maintain the stored information. When power is removed from this memory, its contents are lost.

Non-Volatile Retains its contents when power is removed. This type of memory typically contains calibration or chip configuration information, such as power up states.